



FAR WEST TECHNOLOGY, INC.

330 South Kellogg Ave., Suite D
Goleta, California 93117
USA

Phone: +1 805 964-3615
Fax: +1 805 964-3162
E-mail: info@fwt.com

FWT-70-40M Characterization

Lot 1060

Typical Calibration

(Dose Rate 1.622 kGy/hr)

Dose, Gy	Net A at 600 nm	Net A at 656 nm
30	0.177	
70	0.401	
300	1.644	0.074
700		0.171
1000		0.237
3000		0.631
6000		1.101
10000		1.577

NOTE: This typical calibration is provided as a guide to the response of this lot of FWT-70-40M Opti-Chromic Detectors to ionizing radiation. Actual response also depends on the instrumentation used to measure optical densities.